



Product Service

**Add value.
Inspire trust.**



Executive Summary for the Testing of CORA CHECK-19 swab

Prepared for:

Fameditec GmbH, Düsseldorf

Date: 02.02.2021

Project Number: 713204982

SIGN-ID 405854

02.02.2021

Dr. Christoph Lindner

Dr. Christoph D. Lindner

Team leader Testing Medical Devices

TÜV SÜD Product Service GmbH

Date: see above

SIGN-ID 404124

02.02.2021

Dr. Tobias Beck

Dr. Tobias Beck

Department leader Testing Medical Devices

TÜV SÜD Product Service GmbH

Date: see above

TÜV SÜD Product Service GmbH

Munich Branch
Ridlerstraße 65
80339 Munich
Germany

Phone: +49 89 5008-40
Fax: +49 89 5008-4222

www.tuvsud.com

TÜV®

Page 1



Tests conducted

Test & applicable Standard	Test report summary	Laboratory and test report number
Growth promotion testing acc. to ISO 11737-2:2020	The method suitability test was performed with three test items. Upon results the method is recognized as suitable as no growth inhibitory effect could be detected.	TentaMedix GmbH, Karlsruhe 51958_20210202_VAL_01
Test of sterility acc. to ISO 11737-2:2020	The control complied with acceptance criterion. All five test items were examined as sterile .	TentaMedix GmbH, Karlsruhe 51958_20210202_STE_01

Details regarding the performed tests can be found in the individual test reports.